



# Brake emission measurement

Matteo Federici, Guido Perricone

# Outline



---

- 1. Dyno bench setup;**
- 2. Weighing system;**
- 3. Clean room;**
- 4. Statistical process stability: process control chart;**
- 5. Points to be considered;**



# 1. Dyno bench setup

Inlet air  
(HEPA H13)



Vertical,  
coaxial and  
isokinetic  
sampling

PM<sub>10</sub> cyclons  
flow rates  
10 l min<sup>-1</sup>



Dekati  
DEED

Solid/Total  
PN (CPC and Elpi+)

PM (PM<sub>10</sub> impactor)

Chemical analysis  
(filter holder)



Instrument flow  
rates: 10 l min<sup>-1</sup>  
PM<sub>10</sub> impactor air  
flow controlled with  
a digital mass-flow  
controller



Dekati Elpi+  
10 l min<sup>-1</sup>



TSI 3775  
1.5 l min<sup>-1</sup>



TSI 3790A-10  
1.0 l min<sup>-1</sup>




# 2. Weighing system

Analytical microbalance calibrated according to the DAkkS/DKD regulation

Manufacturer: Sartorius  
Model: MSA2.7S-000-DM



Nominal Weight [mg]	Deviation [mg]	Measuring Uncertainty* [mg]	Relative Uncertainty [%]
10.0000	-0.0002	0.00421	0.0420
15.0000	-0.0001	0.00811	0.0540
100.0000	-0.0001	0.00992	0.0099



Charge neutralizer to stabilize the filter weight

**Measuring range:** 0.0000 – 500.0000 mg


**Readability:** 0.0001 mg (0.1 µg)

**Repeatability:** 0.00005 mg at 200 mg

\*the reported values are determined according to the EURAMET/cg-18/v.04 European guideline.



## 2. Weighing system



- A: rotating tower to stock white and black filters
- B: high precision balance
- C: robotic arm that brings the filters to the balance

## Conditioned + particle-free air glovebox



Environmental conditions:

$22 \pm 1^\circ\text{C}$  & Dew Point  $9.5 \pm 1^\circ\text{C}$  ( $45 \pm 3\%$  RH)

Fully automated procedure, filter conditioning before weighing at least 24 h (*Environmental conditions compliant with ECE Addendum 48: Regulation no. 49*)



# 3.Clean room

The glovebox where the filters are weighed is located into an ISO-6 certified clean room (according to ISO 14644-1)

The preparation of the filters, and the sampling devices (impactor and filter holder) is performed inside the clean room to prevent contaminations




# 4. Statistical process stability: process control charts

19 data points have been collected to generate the process control chart (with the same brake system)

Average value:  $1.92 \pm 0.25^*$   $\text{mg km}^{-1}\text{wheel}^{-1}$

$^* \pm 1\sigma$

Methodology that we developed to check the stability of the measurement over the time



# 5. Points to be considered within TF2

---

- Periodic check of the measurement stability, our proposal is to check the stability every six months.
- Consider that a certain variation in the measured values is expected. What is the allowed variability of the measurement?





# THANK YOU

