

# Inter-Laboratory Comparison Exercise




## CPC CALIBRATION

### CALIBRATION AEROSOLS AND LABORATORY SETUPS

#### Intermediate Status Report

2016-10-12

Alexander Terres



# MOTIVATION

- Comparison of Calibration Aerosols
- Comparison of Laboratory Setups
- Suitability for Calibration sub-23nm ?

# CPC CALIBRATION ROUND ROBIN SCHEDULE

TIME	LABORATORY	Type
02 / 2016	TSI Germany	Instrument Manufacturer
03 / 2016	JRC	Research Institute
04 – 05 / 2016	AVL Austria	Instrument Manufacturer
06 – 07 / 2016	PTB	National Metrological Institute
07 – 08 / 2016	BMW	Vehicle Manufacturer
09 – 10 / 2016	Ricardo Energy & Environment	Calibration Service
11 / 2016	VW	Vehicle Manufacturer
12 / 2016	TSI Germany	Instrument Manufacturer


We are here




# REFERENCE DEVICES & CALIBRATION

LABORATORY	REFERENCE INSTRUMENT	CALIBRATION TYPE
TSI Germany	Aerosol Electrometer	National Metrological Institute
JRC	Aerosol Electrometer CPC	Manufacturer Calibration In-house
AVL Austria	Aerosol Electrometer	Manufacturer Calibration
PTB	Aerosol Electrometer	In-house
BMW Germany	Aerosol Electrometer CPC	Manufacturer Calibration Manufacturer Calibration
Ricardo Energy & Environment	CPC	National Metrological Institute
VW Germany	Aerosol Electrometer CPC	Manufacturer Calibration In-house

# LABORATORY SETUPS




Splitter-type: ISO 27891 compliant setup




Sampling volume-type: not ISO 27891 compliant

# TESTING PROCEDURES: COUNTING EFFICIENCY



Mobility Diameter	10	15	23	41	70	100
Corresponding Diameter, Double Charges	-	-	33	59	103	150
Burner Operating Point	A	A	B	C	D	E
Motivation	Sub-23	Sub-23	PMP	PMP	Linearity	PN-PEMS

# TESTING PROCEDURES: COUNTING EFFICIENCY




# TESTING PROCEDURES: EXEMPLARY EVALUATION

Example: Calibration at 41nm


	Reference Electrometer	Device under Test (CPC)
<b>Raw Concentration</b>	8255/cm <sup>3</sup>	6904/cm <sup>3</sup>
<i>Zero Offset</i>	116/cm <sup>3</sup>	0/cm <sup>3</sup>
Nominal Flow : measured Flow	3 l/min : 2,87 l/min	1 l/min : 1,01 l/min
<i>Flow correction</i>	1,045	0,99
<b>Flow corrected concentration</b>	8629/cm <sup>3</sup>	6836/cm <sup>3</sup>
Flow corr. concentration @59nm	3170/cm <sup>3</sup>	2864/cm <sup>3</sup>
Double charge ratio (Theory)	5,7%	5,7%
Double charged particles	181/cm <sup>3</sup>	163/cm <sup>3</sup>
<i>Double Charge Correction</i>	8629- <b>2*181</b>	6836-163
<b>Charge corrected concentration</b>	8267/cm <sup>3</sup>	6673/cm <sup>3</sup>
<b>CPC Counting Efficiency</b>	6673 : 8267 = <b>80,7%</b>	

# TESTING PROCEDURES: LINEARITY





Reference Device: Aerosol Electrometer or CPC



# TESTING PROCEDURES: COUNTING EFFICIENCY



# SOOT GENERATORS: SIZE RANGE



# SOOT GENERATORS: SIZE RANGE



# SOOT GENERATORS: MAXIMUM CONCENTRATION COMPARISON @10NM



EXAMPLE: Maximum Concentration at BMW laboratory setup (sampling volume, 15l/min dilution air)

	On-site miniCAST		AVL Particle Generator		PALAS DNP 3000	
	GMD	Max. Conc.	GMD	Max. Conc.	GMD	Max. Conc.
10nm	Very small	2500/cm <sup>3</sup>	Very small	35/cm <sup>3</sup>		
	Small	1300/cm <sup>3</sup>	Small	600/cm <sup>3</sup>	Small	200/cm <sup>3</sup>
	Medium	400/cm <sup>3</sup>	Medium	180/cm <sup>3</sup>	Medium	250/cm <sup>3</sup>
15nm	Very small	1700/cm <sup>3</sup>	Very small	25/cm <sup>3</sup>		
	Small	7900/cm <sup>3</sup>	Small	1020/cm <sup>3</sup>	Small	700/cm <sup>3</sup>
	Medium	1700/cm <sup>3</sup>	Medium	1200/cm <sup>3</sup>	Medium	2000/cm <sup>3</sup>

Going to larger sizes (GMD 20nm – 30nm): No need to worry about multiple charges?


→ 10nm: Insufficient concentration for calibration vs. electrometer with example setup

# SOOT GENERATORS: CAST – INTERCHANGING SETTINGS



→ CAST-settings are not transferable between models!


# SOOT GENERATORS: CAST – SINGLE SIZE DISTRIBUTION CALIBRATION



- Size distribution influence
- With **double charge correction** applied
- Especially when comparing Electrometer & CPC

Electrometer Offset  
Very low absolute concentration

- Reference Electrometer
- GMD 20nm
- GMD 35nm
- GMD 50nm



# CPC CALIBRATION ROUND ROBIN

## FIRST FINDINGS

- Participants from all areas of the engine exhaust PN – community
- Measurements examine **CAST soot as a calibration aerosol**
- Possible concentration issues for **sub-23nm calibration**
- No **universal burner settings** for different CAST models
- Size distribution influences measurement results

**Open questions** for calibration with CAST soot  
(only partly covered in ISO 27891):

- Clear definition of suitable **size distributions**
- Accurate correction for **multiple charges** / Strategies to avoid multiple charges
- Zero correction and **Zero drift** of Electrometers
- Suitability of „Sample volume-type“ test setup for calibration

A big “Thank You” to all the labs that are contributing to this Round Robin!

THANK YOU FOR YOUR ATTENTION

